

REVIEW

Can Wearable Devices Accurately Measure Heart Rate Variability? A Systematic Review

Konstantinos Georgiou¹, Andreas V. Larentzakis², Nehal N. Khamis³, Ghadah I. Alsuhaibani³, Yasser A. Alaska³, Elias J. Giallafos⁴

¹ Department of Biological Chemistry, Faculty of Medicine, National and Kapodistrian University of Athens, Athens, Greece

² 1st Propaedeutic Surgical Clinic, Medical School, National and Kapodistrian University of Athens, Athens, Greece

³ Clinical Skills & Simulation Center, Department of Medical Education, College of Medicine, King Saud University, Riyadh, KSA

⁴ Medical Physics Simulation Center, Medical School, National and Kapodistrian University of Athens, Athens, Greece

Correspondence:

Konstantinos Georgiou, Department of Biological Chemistry, Faculty of Medicine, National and Kapodistrian University of Athens, 75 Mikras Asias Str., Goudi, 11527 Athens, Greece
E-mail: kongeorgiou@med.uoa.gr
Tel: +306942066216

Received: 27 Dec 2017

Accepted: 19 Jan 2018

Published Online: 31 Jan 2018

Published: 30 March 2018

Key words: heart rate, heart rate variability, wearable, photoplethysmography

Citation: Georgiou K, Larentzakis AV, Khamis NN, Alsuhaibani GI, Alaska YA, Giallafos EJ. Can wearable devices accurately measure heart rate variability? A systematic review. *Folia Med (Plovdiv)* 2018;60(1):7-20.

doi: 10.2478/folmed-2018-0012

Background: A growing number of wearable devices claim to provide accurate, cheap and easily applicable heart rate variability (HRV) indices. This is mainly accomplished by using wearable photoplethysmography (PPG) and/or electrocardiography (ECG), through simple and non-invasive techniques, as a substitute of the gold standard RR interval estimation through electrocardiogram. Although the agreement between pulse rate variability (PRV) and HRV has been evaluated in the literature, the reported results are still inconclusive especially when using wearable devices.

Aim: The purpose of this systematic review is to investigate if wearable devices provide a reliable and precise measurement of classic HRV parameters in rest as well as during exercise.

Materials and methods: A search strategy was implemented to retrieve relevant articles from MEDLINE and SCOPUS databases, as well as, through internet search. The 308 articles retrieved were reviewed for further evaluation according to the predetermined inclusion/exclusion criteria.

Results: Eighteen studies were included. Sixteen of them integrated ECG - HRV technology and two of them PPG - PRV technology. All of them examined wearable devices accuracy in RV detection during rest, while only eight of them during exercise. The correlation between classic ECG derived HRV and the wearable RV ranged from very good to excellent during rest, yet it declined progressively as exercise level increased.

Conclusions: Wearable devices may provide a promising alternative solution for measuring RV. However, more robust studies in non-stationary conditions are needed using appropriate methodology in terms of number of subjects involved, acquisition and analysis techniques implied.

BACKGROUND

Monitoring and analysis of heart rate (HR) provide valuable information regarding health status and have been extensively investigated in various activities of healthy subjects as well as in patients suffering from various diseases.^{1,2} Heart rate variability (HRV) has emerged as a non-invasive tool to estimate the vagal activity in several conditions, including monitoring of athletic responses to training. Decreased HRV has been reported as a predictive factor for adverse outcomes in disease states and has been found to be associated with fatigue, stress, and even burnout during athletic performance.³⁻⁵

High HRV is an indication of a better general health status as it allows better adjustment to external and internal stimuli.⁶

Due to the fact that traditional HRV recording methods, such as using electrocardiography (ECG) and specialized software, often involve expensive equipment, which is primarily found in research laboratories, alternative methods have been used, yet, with variable results. Photoplethysmography (PPG) is a simple and low-cost method used to detect volumetric changes in the peripheral blood circulation at the skin surface.⁷

In recent years, several wearable pulse rate moni-

tors using PPG technology have been developed and have become widely available. The general concept of operation of these small, robust and user-friendly devices is that they contain sensors which reliably monitor minor changes in the intensity of light from high intensity light emitting diodes (LEDs) that is transmitted through or reflected from the human tissues. Although they have obvious advantages over the classical ambulatory ECG recording, the fact that they use PPG, i.e. a different detecting approach, raises the question of how much accurate and reliable are their results when compared to the gold standard ECG method?

The purpose of this systematic review is to present the available literature comparing the ECG derived RR and HRV with that of the wearable commercially available devices in terms of accuracy and reliability as well as to reveal their strengths and limitations in the everyday clinical practice.

MATERIALS AND METHODS

SEARCH STRATEGY

This systematic review was conducted by searching medical literature in MEDLINE and SCOPUS. The search was guided by the Preferred Reporting Items for Systematic Reviews and Meta-Analyses (PRISMA) statement in conjunction with the PRISMA explanation and elaboration document.^{8,9} The last search was conducted in April 2017. All the retrieved article titles and abstracts were screened for relevant manuscripts. Full texts were then retrieved for the relevant articles and these were thoroughly read to judge if they meet the inclusion criteria of the current systematic review.

Medical Subject Heading (MeSH) terms and text words were used based on the following search strategy:

Group A terms: (HRV) OR heart rate variability

Group B terms: (((smart) OR Smartphone\$) OR wearable\$) OR phone\$ OR plethysmography OR photoplethysmography OR impedance)

Group C terms: (((holter) OR continuous ECG) OR continuous electrocardiogram) OR continuous electrocardiography) OR ambulatory ECG) OR ambulatory electrocardiogram OR ambulatory electrocardiography.

Group A, group B, and group C were combined and humans' studies and English language limits were applied.

We also searched the reference lists of retrieved full text review manuscripts for the relevant articles (regardless of their publishing date) and included

these in our systematic review.

Well defined inclusion/exclusion criteria were applied to filter the retrieved literature. Of the articles retrieved through the above described search strategy only those that met the following criteria were considered for this systematic review:

Inclusion criteria:

Studies related to human subjects.

Studies reported in full text English language.

Studies related to HRV. Studies on HRV detection and/or analysis and/or interpretation and/or filtering were included.

Studies that compare ECG/Holter recordings of HRV with any other wearable HRV detection and capture method(s) were included.

Original papers were only included; however, reference lists from other kind of manuscripts were used to retrieve any relevant original studies.

Exclusion criteria:

Studies that did not include HRV were excluded.

Studies that did not use commercially available wearable hardware were excluded.

In addition, we performed Google search for wearable devices which claimed that they can measure HRV via plethysmography HR and are available on the market. Our search spanned the last 5 years. The devices list (**Table 1**) was used as an additional retrograde search tool for any relevant studies through manufacturers' commercial websites.

Two of the authors (KEG and AL) independently applied the above described search strategy to retrieve and screen the articles. Any disagreements were resolved by a third author (AJG) and a final decision was made accordingly.

RESULTS

Applying the search strategy described in the methods section, we retrieved 57 articles from MEDLINE and 269 from SCOPUS. Exclusion of duplicates yielded 308 articles. Of these articles, 272 were excluded after title and abstract screening according to the predefined criteria. Thirty-six articles were selected for full text review. Full text reading resulted in exclusion of further 30 articles, which were found to be either irrelevant (n=28) or non-comparison papers (n=2). The remaining six articles were included in the study. Additionally, 12 relevant peer-reviewed articles were identified from the reference list of the reviews and from manufacturers' commercial websites and added to the study. So, 18 articles were finally included in this systematic review. The flow diagram of the selection process is shown in **Fig. 1**.

Table 1. Capable to measure HRV wearable models: sensor location, release year and connectivity type

	Model	Sensor site	Company	Release Year	Connectivity
1	4III VIIIIVA	chest	4III	2013	B & A
2	60beat HR Monitor	chest	60beat	2016	B & A
3	HRM Blue	chest	BlueLeza	2016	B & A
4	Dash Earphones	ear	Bragi	2014	B
5	TP3	chest	Cardiosport	2015	B & A
6	Hexoskin	torso	Carre Technologies	2014	B & USB
7	Empatica E4 Wristband	wrist	Empatica	2016	B
8	R2 Smart Fitness HRM Wristband	chest	Cositea	2016	B & A
9	EQ02 LifeMonitor	chest	Equivital	2012	B, P & USB
10	Forerunner 935	wrist	Garmin	2016	B & A
11	HRM Tri	wrist	Garmin	2016	B & A
12	Premium Heart Rate Monitor	chest	Garmin	2016	B & A
13	Vivoactive HR+	wrist	Garmin	2015	B & A
14	910XT	wrist	Garmin	2016	B, P & A
15	920XT	wrist	Garmin	2016	B, P & A
16	HeartMath emWave Pro	ear	HeartMath	2015	B & A
17	Athos	torso, legs & thighs	Mad Apparel	2015	B
18	HxM Smart HR	chest	Medronic	2017	B & Gateway
19	Alpha 2	wrist	Mio	2015	B & A
20	Oxstren	hand	Oxstren	2015	B
21	H7 Heart Rate Monitor	chest	Polar	2012	B & WiFi
22	H10	chest	Polar	2016	B & WiFi
23	QardioCore	chest	Qardio	2017	B
24	SmartBand 2	wrist	Sony	2016	B
25	Smart Sensor	chest	Suunto	2017	B & A
26	Spartan Sport	wrist	Suunto	2017	B & A
27	DGYAO® Mobile HR Monitor	ear	Top Yao	2016	B
28	Tickr Heart Rate Monitor	chest	Wahoo Fitness	2015	B & A
29	Tickr X Workout Tracker	chest	Wahoo Fitness	2017	B & A
30	WHOOP Strap 2.0	wrist	Whoop	2016	B
31	Tinké Fitness & Wellness Tracker	fingers	Zensorium	2016	B & A

B: Bluetooth, A: ANT+, P: proprietary




Figure 1. Flow diagram.

Table 2 summarizes the quantitative and qualitative methodology used and the categorization (measurement, position, location, sample number, technology used) as well as the main findings of the included studies.

The total number of subjects involved in the included studies was 686, with one study having 339 subjects.¹⁰ All 18 articles have examined healthy subjects except one: Vasconcellos et al. (2015) who studied obese adolescents.¹¹ Also, one study examined children.¹² In most of the articles (n=16), a chest device was used,¹⁰⁻²⁵ while a finger device was used in two studies^{26,27}. Sixteen of the studies utilized a similar to ECG - HRV technology¹⁰⁻²⁵ and the remaining two integrated

PPG - PRV technologies^{26,27}.

REST

All studies examined the HRV/PRV at rest as baseline. Overall, agreement between the several indices of HRV and PPG as measured by Holter and wearable devices, respectively, was very good to excellent (ranging from 0.85 to 0.99). The RR interval correlation ranged from 0.91 to 0.999. Moreover, in two studies,^{12,15} the error rate in detection of R waves was evaluated and found to range between 0.28 and 0.4%. This was estimated as an accepted ratio. Regarding time domain indices of HRV, correlations ranged from 0.98 to 0.99, while in the frequency domain the correlation was

Table 2. Results

#	Study	Rest	Supine	Upright / Sitting	Walking / Running / Exercise	Wearable device Location	n	Wearable device Technology	Comparison	Details
1	Akintola et al., 2016	Y	Y	Y	Y	Chest (Belt)	18	ECG	Correlation depends on artifacts	Average artifact % → 19%. AMV (r: 0.967), SDNN (0.393), RMSSD (0.285), SDANN (0.680), pnn50 (0.982).
2	Esco & Flatt 2017	Y	Y	Y	N	Finger (silicone sheath)	30	PPG	Good agreement	LOA Ln RMSSD constant error ± SD: -0.13 ± 2.83 for the supine values, -0.94 ± 3.47 for the seated values, -1.37 ± 3.56 for the standing values. (r values from 0.98 to 0.99).
3	Flatt & Esco, 2013	Y	Y	N	N	Chest (Belt)	25	ECG	Total agreement	No significant difference, correlation nearly perfect for RMSSD (r: 0.99).
4	Gamelin et al, 2006	Y	Y	Y	N	Chest (Belt)	20	ECG	Good agreement	Supine vs. standing: differences for uncorrected & corrected RR coefficient correlation: 0.88 & 0.91 for supine & standing. No differences except RMS-SD, SD ₁ in standing. Detection Error Rate of R waves: 0.4%.
5	Gamelin et al, 2008	Y	Y	N	N	Chest (Belt)	12	ECG	No difference	Correlation between ECG & Polar RR intervals (corrected & uncorrected) was 0.80. No significant differences for Time Domain, FFT & Poincare plot except for SD ₂ . R waves detection error rate: 0.28%.

6	Giles et al, 2016	Y	Y	Y	N	Chest (Belt)	18	ECG	Accurate. Good agreement	No significant differences for SDNN, RMSSD, pNN50, VLF, LF, HF, nnLF
7	Heathers, 2013	Y	N	Y	Y	Finger (silicone sheath)	20	PPG	Good agreement at rest. Mildly reduced agreement during exercise.	Experiment 1: Close agreement – small overall bias & acceptable limits of agreement. RR & PP correlation coefficient (0.988 – 0.999). All HRV from SPRV > than ECG. Experiment 2: Close agreement between RR & PP (highest at rest, r: 0.993 – 0.997, slightly attenuated in exercise (0.965 – 0.998).
8	Hernando et al, 2016	Y	N	Y	Y	Chest (Belt)	23	ECG	As exercise increases, correlation decreases	19/23 had high correlation at rest ($r^2 > 0.8$). Discrepancy increased from 1.67% (at rest) → 4.8% at the exercise peak. As exercise increases, reliability & agreement indices drop below 0.5.
9	Hong et al, 2009	N	N	N	Y	Chest (Bioshirt)	18	ECG	As exercise increases, correlation decreases (artefacts)	Coefficient correlation of HRV (r^2 : 0.965 or higher), HF, LF
10	Kingsley et al, 2005	Y	N	Y	Y	Chest (Belt)	8	ECG	High agreement higher at rest. Reduced agreement at exercise.	Short relationship between RR internals during exercise. Similar results obtained for the RR internal but significant differences occurred for HRV indices.
11	Nunan et al, 2008	Y	Y	N	N	Chest (Belt)	33	ECG	Accurate / Good agreement	No significant differences for HRV indices – LOA for mean RR, LFnn, HFnn, LF/HF
12	Nunan et al, 2009	Y	Y	N	N	Chest (Belt)	33	ECG	Good agreement Not for all indicators	Correlation coefficient: 0.99, 0.86, 0.85 for mean RR, LFnn, HFnn respectively. Near perfect correlation for SDNN, RMSSD (0.99, 0.37). Good correlation for LF (0.92), HF (0.94), LF/HF (0.87). All measures of HRV ranged from 0.85-0.99.

13	Plews et al, 2017	Y						Chest (Belt)	29	ECG	Acceptable agreement	Almost perfect correlation all HRV indices (RMSSD)
14	Romagnoli et al, 2014	Y	N	Y	Y			Chest (Vest)	12	ECG	Good agreement for RR / unequal results for HRV.	Mean RR, SDNN & SD ₂ , excellent LOA. RMSSD, HF, LF/HF, HF & SD ₁ worst LOA.
15	Vanderlei et al, 2008	Y	N	Y	Y			Chest (Belt)	15	ECG	As exercise increases, correlation decreases.	HRV indices assessed: PNN50, RMSSD, LFNu, HFNu, LF/HF.
16	Vasconcellos et al, 2015	Y	Y					Chest (Belt)	14	ECG	Very good correlation	Moderate to strong agreement (0.68-0.98) for HR, RR, pNN50, rMSSD, LF, HF, LF/HF. In general moderate agreement for frequencies domain measures.
17	Wallen et al, 2012	Y	Y	N	N			Chest (Belt)	339	ECG	Good to excellent agreement	328 subjects. Gender & age dependent: females especially > 60 yrs old present reduced correlation). ICC coefficients ↓ in women than men → more pronounced in SDNN.HRV from Polar → Gender + Age dependent. ICC > 0.8 on all HRV parameters among men. Agreement moderate on all HRV among men.
18	Weippert et al, 2010	Y	Y	N	N			Chest (Belt)	19	ECG	Limited agreement	Good correlation (narrow LOA & small bias). ICC for HRV frequencies were high, however in most cases LOA showed unacceptable discrepancies.

LOA: level of agreement

found to range from 0.85 to 0.94. Finally, in the non-linear multivariate analysis, the correlation was found to be > 0.9 .

EXERCISE

Eight out of the 18 examined studies used an exercise protocol, with a total of 131 participants. All of them reported that although there was an excellent correlation between HRV and PPG as measured by Holter and wearable devices at rest, this seemed to decrease up to 0.85 as the level of exercise and/or motion increased. Overall, RR agreement was moderate to excellent ranging from 0.786 to 1. Regarding time domain, HRV parameters correlation was found to range from 0.786 (at the exercise peak) to 1 (rest and exercise 1st phase). Similar pattern occurred in the frequency domain HRV parameters where the correlation was found to range from 0.8 to 1. Also, RR to pulse pressure (PP) wave correlation was ranging from 0.8 to 0.998.

DISCUSSION

In recent years, technology advances have been used to capture HRV through wearable devices during daily activities. The accuracy of these devices versus classical methods like ECG is still under evaluation.

This systematic review aims to present the available relevant literature and discuss their findings, their limitations as well as to provide possible explanations for these findings.

It is important to realize that the basic difference between PPG and ECG is the captured signal per se: the electrical activity of the heart is depicted by ECG, whereas the PPG is a mechanical signal measuring the propagation of the peripheral pulse wave. Therefore, the time of propagation of the PP wave from the heart to the distal arterioles is called pulse transit time (PTT). It is a measure of the time that elapsed between the R-wave of QRS complex in the ECG and the arrival point to PPG device.²⁸ Several studies have shown that PTT seems to be a surrogate marker of ANS in parallel to HRV²⁹ and that PTT is dependent on the properties of the pulse wave velocity, the vascular path from the heart to the location of the detector and is negatively correlated with blood pressure, arterial stiffness and age.⁷

Therefore, we will discuss HRV and PPG fundamentals as well as the factors affecting them in order to elaborate PPG versus ECG for HRV measurement.

HRV FUNDAMENTALS

HRV analysis is a widely available and accurate non-invasive technique used as a quantitative assessment tool of the autonomic nervous system (ANS) function.³⁰ Studies have shown that reduced HRV indices, as assessed by the RR interval analysis, is associated with increased cardiovascular morbidity and mortality in patients with various diseases as well as in the general population.³¹ In addition, heart rate and HRV analysis have been used for estimation of mental stress and athlete's fitness levels, fatigue and overload.⁵

The quantification of ANS function is feasible by calculating several HRV parameters according to time-domain, frequency-domain and nonlinear analysis of consecutive RR intervals.¹ These indices represent different components of the sympathetic and/or parasympathetic system of the ANS. For instance, the high frequency (HF) component derived by the frequency domain analysis represents the parasympathetic activity, while the LF/HF ratio represents the balance of sympathetic to parasympathetic activity.³²

However, it is important to realize that these HRV indices depends on the recording quality, the subject's activity during the recording, the exclusion of artifacts, the detection of arrhythmic beats and the recording duration ranging from seconds (short term HRV) to even days (long term). Some of these indices, like the root mean square of standard deviation (RMSSD) of RR interval can be calculated from a 10-second session while others need more than one-hour recording time.³³

Despite being accepted as gold standard methods for RR interval monitoring and analysis, both the classical ECG and the ambulatory Holter monitoring have several drawbacks regarding the proper and accurate detection of RR intervals. For example, patients with tremor or elderly patients with fragile skin have bad quality of recordings with a lot of noise and artifacts.³⁴ Similarly, other factors such as surface electromyography, increased electrode impedance, respiration induced baseline drift, and electrode contact movement can cause noise and motion artifacts.⁷ Additionally, morphological variations in the ECG waveform and heterogeneity in the QRS complex can often make difficult to identify the RR interval.³⁵ Another limitation can be the need for the presence of a specialized technician/doctor, thus increasing the cost and accordingly decreasing the wide applicability. Finally, a reported drawback in ECG wearable devices that do not record standard ECG derivations is their inability

to distinguish some arrhythmias and ectopic beats.¹⁰

PPG FUNDAMENTALS

Photoplethysmography (PPG), a cheap, simple and widespread technology has been used as an alternative approach to obtain HRV indices.⁷ The PPG based devices have a sensor that uses infrared emitter and a detector. This emitter is integrated to a probe which is comfortable to wear in stable places of the body that are rich in microcirculation. Thus, the blood volume changes in the microvascular bed which are synchronous to the heartbeat can be traced without the inconveniences of electrode installation or the need to undress the examinee.¹ The simplicity of the technique, cost-effectiveness, easy signal acquisition and remote monitoring are the main and obvious advantages of the PPG versus the gold standard ECG. Therefore, PPG is often used in conditions of measurements where mobility, simplicity, time efficiency, flexibility and low cost are of paramount importance, e.g., in exercise monitoring, every day motion, monitoring of the elderly, or with disabled patients, etc.^{7,36}

In the relevant literature of the wearable PPG devices the terms 'heart rate' and 'pulse rate' are frequently used interchangeably. Also, the term 'pulse rate variability' (PRV), which is derived from PPG, has been suggested as a potential analog of HRV.^{37,38}

As PRV is further processed identically to HRV, the derived parameters can be extracted from both methods such as the standard deviation of normal to normal (SDNN) R-R intervals (NN), root mean square of successive differences between adjacent NN intervals (RMSSD), proportion of NN50 in total NN intervals (pNN50), low frequency (LF) power, high frequency (HF) power and LF/HF ratio (LF/HF).

Recently, latest technology smart phones applications with wearable devices^{26,27} use PPG for assessing HRV as an alternative to ECG monitoring in clinical research. More information is available in **Table 2**.

WAVELENGTH USED

PPG uses low intensity infrared or green light, which are more strongly absorbed by the blood than the surrounding tissues. It has been shown that 530 nm light (green) PPG showed higher accuracy of pulse rate detection than the 645 (red) and 470 (blue) nm light for monitoring HR.³⁹

MEASUREMENT SITE

A flat skin surface with rich microvasculature is

required to firmly attach the PPG sensor to obtain an accurate measurement. As such, the usual measurement sites for wearable PPGs are the wrist and the chest. Most of the wearable devices are placed on the wrist and considerably fewer of them on the chest.

However, it is worth noting that there is also a bunch of quasi-wearable devices attached to either the ear lobe or the finger which compared HRV to PRV parameters.

The ear is chosen as a measuring site because it is a natural anchoring point, and it is discrete since the device can be partially hidden by hair. Weinschenk et al.²⁸ compared PPG to ECG HRV measurements in deep breathing test in 343 female subjects in resting conditions and using appropriate parameters they found an excellent correlation. Finger tips have also been used as a measuring site but only in stationary conditions: it has been shown that HRV derived from fingertip PPG had an excellent correlation to ECG in stationary conditions.³⁸ However, a comparative study of finger derived PRV and HRV in healthy subjects⁴⁰ using a stationary equipment found a poor correlation and suggested that finger derived PRV might not be suitable to substitute ECG derived HRV, as has been mentioned by other authors, too⁴¹.

The use of a smart phone camera as PPG sensor can provide an acceptable agreement for some HRV indices when compared with ECG,²² but there are increasing differences in HRV and HR detection in the setting of movements or exercise⁴². This is because the application algorithm is cancelled in the filtering process of inaccurate signals as well as due to the difficulty of stabilizing the fingertip on the camera while exercising.

PARAMETERS INVOLVED IN PPG MEASUREMENTS

Several parameters have to be considered when interpreting PPG measurements. These include:

1. *Motion artifact*: Special attention must be exercised during data PPG acquisition to eliminate motion-induced artifacts.⁷ The contact force between the site and the sensor should be considered as PPG is vulnerable to such type of artifacts. However, despite the importance of this factor, we found only one study in 16 male ischemic patients measuring the accuracy of a smart phone derived pulse rate versus ECG, where an excellent correlation was found at rest, which was slightly deteriorated during exercise.⁴²
2. *Respiration*: Since respiration alters the intrathoracic pressure and causes blood flow variations

in both the veins the DC component of the PPG waveform shows minor changes with respiration.⁴³ Thus, it has been shown that the short-term variability (RMSSD, SD1, and HF) and LF/HF agreement between PRV and HRV deteriorates as a result of the vulnerability to respiratory changes.⁴⁴

3. *Age, gender and environmental factors*: Normal HRV values for various age and gender groups are still not available in the literature. However, it is well known that the elderly have increased arterial stiffness which leads to faster pulse transmission in the periphery and thus pulse transit time (PTT) differences observed between HRV and PRV could be attributed to aging.⁴⁵ The reviewed articles, except one,²³ involved young population (mean range from 20.9 to 39.2 years).

Regarding gender influence, in our review, just one study showed that measuring HRV at rest was age and gender dependent, the correlation with ECG being lower in women than men and further decreasing in older women.¹⁰ As there is no strong evidence provided so far that age and/or gender can play a role in the studied correlations, further studies are needed to investigate these two variables in different populations while using appropriate sampling and prospective study design with a longitudinal follow-up.²⁸

Environmental factor effects such as temperature was investigated in one of the studies. This study concluded that ambient temperature could induce a difference in the short-term variables that reflect the parasympathetic activity between HRV and PRV.³⁸

4. *Software analysis*: Some proprietary software systems for collection and analysis of RV data exist like the PPT5 or the Ithlete™ software application^{14,26,27} or freely available software may be used e.g. Kubios (<http://www.kubios.com>)⁴⁶. According to guidelines², manual editing should be preferred instead of automated data analysis as automatic filters are known to be unreliable and may potentially introduce errors. In our review only six studies used automated analysis only whether the rest used both manual and automated analysis.

5. *Statistical analysis*: The Bland-Altman plot must be used to compare the agreement among a new measurement technique with a gold standard, as even a gold standard does not imply to be without error. This plot allows the identification of any systematic difference between the measurements.⁴⁷

In our review only 17 studies used this technique while four studies did not apply the Bland-Altman analysis and therefore only the correlation, but

not agreement between the two methods, could be determined from these publications.

6. *Sampling rate*: The sampling rate is a matter of difference between the two approaches. Sampling rate of PPG is usually 20 Hz much less than that of ECG which is 125 to 250 Hz. This obviously implies weaker ability of the PPG devices for events detection.³⁶

COMPARISON OF PPG VS. ECG FOR HRV MEASUREMENT

There are several studies examining the correlation between HRV and PRV with inconclusive results.⁴⁰ This may be due to different experimental settings or to the absence of standardization of the methods of analysis used.⁴⁴ It is worth noting that the disagreement between the two methods does not apply to the same extent to all HRV parameters. Additionally, PPG is susceptible to motion artifacts. As such, the accuracy of PRV as obtained from PPG should be interpreted with caution.³⁶

REST

Our search revealed that the comparison studies performed in stationary conditions have generally revealed that PRV is a good surrogate of HRV. This is in line with other studies not involving wearable devices, which also found an acceptable agreement between HRV and PRV at sitting and resting positions.⁴⁸

PRV becomes stronger at a standing than at a supine position, as it reflects the mechanical coupling between respiration and thoracic vasculature tone. Therefore, when a subject changes his/her position from supine to upright, even in resting conditions, a PRV divergence from HRV becomes apparent. Additionally, HRV indices derived from PPG data are very sensitive to different factors including noise, artifacts, stature, atherosclerosis, location of sensor and sampling rate. It is probably due to these reasons that some studies comparing the two methods found differences among normal healthy subjects^{7,36,45} as well as in patients⁴⁰.

EXERCISE

There are many non-stationary situations where autonomic balance significantly changes like in stress, during motion or exercise. Unfortunately, in such situations where PRV would be more useful as a surrogate measurement of HRV, its clinical value is questionable: a moderate agreement was observed in some studies about the factors affecting the measurements when the subject is exercising or having mental stress i.e. increased noise produc-

tion, contraction of muscles which are in contact with the sensors, sweating, increased intrathoracic pressure altering the venous return, increased peripheral vasoconstriction and the respiratory effort during exercise.

RR interval variables of Bioshirt ECG were compared to those from conventional ECG and found that R-peak detecting capabilities of these two devices were largely similar. However, as the level of exercise was increasing, the correlation was decreasing due to artifact production. It must also be noted that a disadvantage of chest band wearable devices during intense exercise is the discomfort that a subject senses as the chest expands with deep breathing.¹⁸

Hernando et al. (2016)¹⁷ observed that although an agreement between the detected R-peaks and the RR intervals from the Polar wearable and ECG existed, as the exercise intensity level was increased, the discrepancy of the RR pairs Bland-Altman plot also increased. They noticed a good correlation in some but not in all of the HRV indices, due to the disagreement of the relative error of the Polar derived high frequency with that of the ECG, as the level of the exercise increases.¹⁷

Akintola et al. (2016) used a chest wearable device detecting ECG and HRV and reported enormous amount of artifacts during daily activities in 18 healthy subjects and they concluded that this is a limitation of the wearable device used.¹³

In agreement with all of the above, another study showed that the limits of agreement were deteriorating as the exercise was intensified, implying an influence of adrenergic input, respiratory effort and unreliable algorithm detection and recording RR ability.¹⁹

In contrast to the many negative results reported during intense exercise, there are other researchers who reported an overall stronger agreement.⁴⁹ Also, some other research groups reported that PPG yielded higher HRV values.³⁶ However, all these studies involved only a sample size of few subjects.

Unfortunately, most of the findings from our review showed that the correlation was fading out as the level of exercise and/ or motion increases. Furthermore, the data from the reviewed studies are not able to support an in-depth quantitative analysis due to the differences in their methodology.

FUTURE DIRECTIONS

As wearable healthcare technology and the research of light propagation in human tissues are progressing,

it is expected that PPG applications will expand. For instance, there is a growing interest to remotely depict PPG through imaging such as contactless video-photoplethysmography (vPPG)⁵⁰ or imaging PPG (IPPG)⁵¹.

It is essential to develop advanced wearable devices with higher accuracy, to minimize motion artifacts as well as improved algorithms to better detect and identify errors that may occur during exercise and higher intensity motion.

Additionally, as the availability of wearable devices is expanding, more research is obviously warranted to establish age- and sex-dependent normal PRV values as well as to standardize both acquisition protocols and analytical methods in order to get reliable and accurate results, thus permitting these methods to become a valid surrogate for HRV parameters.

CONCLUSION

Our systematic review revealed that wearable devices, especially those using PPG, may provide a promising alternative solution for measuring HRV. However, it is evident that more robust studies in non-stationary conditions are needed with appropriate methodology in terms of number of subjects involved, acquisition and analysis techniques implied, before being able to recommend any of the commercially available devices. Therefore, so far wearable devices can only be used as a surrogate for HRV at resting or mild exercise conditions, as their accuracy fades out with increasing exercise load.

REFERENCES

1. Liu AB, Wu HT, Liu CC, et al. The factors influence compatibility of pulse-pulse intervals with R-R intervals. In: EMBC 2013. Proceedings of 2013 Annual International Conference of the IEEE: Engineering in Medicine and Biology Society Piscataway, NJ: IEEE Service Center. 2013; p. 2068-71.
2. Task force of the European Society of Cardiology and the North American Society of Pacing and Electrophysiology. Heart rate variability: Standards of measurement, physiological interpretation and clinical use. *Circulation* 1996;93(5):1043-65.
3. Buccelletti E, Gilardi E, Scaini E, et al. Heart rate variability and myocardial infarction: systematic literature review and meta-analysis. *Eur Rev Med Pharmacol Sci* 2009;13(4):299-307.
4. La Rovere MT, Pinna GD, Maestri R, et al. Autonomic markers and cardiovascular and arrhythmic events in heart failure patients: still a place in prognostication? Data from the GISSI-HF Trial. *Eur J*

- Heart Fail 2012;14(12):1410-9.
5. von Rosenberg W, Chanwimalueang T, Adjei T, et al. Resolving ambiguities in the LF/HF ratio: LF-HF scatter plots for the categorization of mental and physical stress from HRV. *Front Physiol* 2017;8:360.
 6. Lehrer PM, Vaschillo E, Vaschillo B, et al. Heart rate variability biofeedback increases baroreflex gain and peak expiratory flow. *Psychosom Med* 2003;65(5):796-805.
 7. Allen J. Photoplethysmography and its application in clinical physiological measurement. *Physiol Meas* 2007;28:R1-39.
 8. Moher D, Liberati A, Tetzlaff J, et al. Preferred reporting items for systematic reviews and meta-analyses: the PRISMA statement. *J Clin Epidemiol* 2009;62(10):1006-12.
 9. Liberati A, Altman DG, Tetzlaff J, et al. The PRISMA statement for reporting systematic reviews and meta-analyses of studies that evaluate health care interventions: explanation and elaboration. *J Clin Epidemiol* 2009;62:e1-34.
 10. Wallén MB, Hasson D, Theorell T, et al. Possibilities and limitations of the Polar RS800 in measuring heart rate variability at rest. *Eur J Appl Physiol* 2012;112(3):1153-65.
 11. Vasconcellos FV, Seabra A, Cunha FA, et al. Heart rate variability assessment with fingertip photoplethysmography and Polar RS800cx as compared with electrocardiography in obese adolescents. *Blood Press Monit* 2015;20(6):351-60.
 12. Gamelin FX, Baquet G, Berthoin S, et al. Validity of the Polar S810 to measure R-R intervals in children. *Int J Sports Med* 2008;29:134-8.
 13. Akintola A, van de Pol V, Bimmel D, et al. Comparative analysis of the Equivital EQ02 life monitor with holter ambulatory ECG device for continuous measurement of ECG, heart rate, and heart rate variability: A validation study for precision and accuracy. *Front Physiol* 2016;7:391.
 14. Flatt A, Esco MR. Validity of the Ithlete™ smart phone application for determining ultra-short-term heart rate variability. *J Hum Kinet* 2013;39:85-92.
 15. Gamelin FX, Berthoin S, Bosquet L. Validity of the Polar S810 heart rate monitor to measure R-R intervals at rest. *Med Sci Sports Exerc* 2006;38:887-93.
 16. Giles D, Draper N, Neil W. Validity of the Polar V800 heart rate monitor to measure RR intervals at rest. *Eur J Appl Physiol* 2016;116:563-71.
 17. Hernando D, Garatachea N, Almeida R, et al. Validation of heart rate monitor Polar RS800 for heart rate variability analysis during exercise. *J Strength Cond Res* 2016; doi: 10.1519/JSC.0000000000001662.
 18. Hong S, Yang Y, Kim S, et al. Performance study of the wearable one-lead wireless electrocardiographic monitoring system. *Telemed J E Health* 2009;15:166-75.
 19. Kingsley M, Lewis MJ, Marson RE. Comparison of Polar 810s and an ambulatory ECG system for RR interval measurement during progressive exercise. *Int J Sports Med* 2005;26:39-44.
 20. Nunan D, Jakovljevic DG, Donovan G, et al. Levels of agreement for RR intervals and short-term heart rate variability obtained from the Polar S810 and an alternative system. *Eur J Appl Physiol* 2008;103:529-37.
 21. Nunan D, Donovan G, Jakovljevic DG, et al. Validity and reliability of short-term heart-rate variability from the Polar S810. *Med Sci Sports Exerc* 2009;41:243-50.
 22. Plews DJ, Scott B, Altini M, et al. Comparison of heart rate variability recording with smart phone photoplethysmographic, Polar H7 chest strap and electrocardiogram methods. *Int J Sports Physiol Perform* 2017;12(10):1324-8.
 23. Romagnoli M, Alis R, Guillen J, et al. A novel device based on smart textile to control heart's activity during exercise. *Australas Phys Eng Sci Med* 2014;37:377-84.
 24. Vanderlei LC, Silva RA, Pastre CM, et al. Comparison of the Polar S810i monitor and the ECG for the analysis of heart rate variability in the time and frequency domains. *Braz J Med Biol Res* 2008;41:854-9.
 25. Weippert M, Kumar M, Kreuzfeld S, et al. Comparison of three mobile devices for measuring R-R intervals and heart rate variability: Polar S810i, Suunto t6 and an ambulatory ECG system. *Eur J Appl Physiol* 2010;109:779-86.
 26. Esco MR, Flatt AA, Nakamura FY. Agreement between a smartphone pulse sensor application and electrocardiography for determining lnRMSSD. *J Strength Cond Res* 2017;31:380-5.
 27. Heathers JA. Smartphone-enabled pulse rate variability: an alternative methodology for the collection of heart rate variability in psychophysiological research. *Int J Psychophysiol* 2013;89:297-304.
 28. Weinschenk SW, Beise RD, Lorenz, J. Heart rate variability (HRV) in deep breathing tests and 5-min short-term recordings: agreement of ear photoplethysmography with ECG measurements, in 343 subjects. *Eur J Appl Physiol* 2016;116:1527-35.
 29. Drinnan MJ, Allen J, Murray A. Relation between heart rate and pulse transit time during paced respiration. *Physiol Meas* 2001;22(3):425-32.
 30. Gisselman AS, D'Amico M, Smoliga JM. Optimizing inter-session reliability of heart rate variability - the effects of artifact correction and breathing type. *J Strength Cond Res* 2017. doi: 10.1519/

- JSC.0000000000002258.
31. Maheshwari A, Norby FL, Soliman EZ, et al. Low heart rate variability in a 2-minute electrocardiogram recording is associated with an increased risk of sudden cardiac death in the general population: the atherosclerosis risk in communities study. *PLoS One* 2016;11(8):e0161648.
 32. Georgiou K, Larentzakis A, Papavassiliou A. Surgeons and surgical trainees acute stress in real operations or simulation: a systematic review. *Surgeon* 2017;15(6):355-65.
 33. Shaffer F, Ginsberg JP. An overview of heart rate variability metrics and norms. *Front Public Health* 2017;5:258.
 34. Pérez-Riera AR, Barbosa-Barros R, Daminello-Raimundo R, et al. Main artifacts in electrocardiography. *Ann Noninvasive Electrocardiol* 2017; doi:10.1111/anec.12494.
 35. Gambarotta N, Aletti F, Baselli G, et al. A review of methods for the signal quality assessment to improve reliability of heart rate and blood pressures derived parameters. *Med Biol Eng Comput* 2016;54:1025-35.
 36. Schäfer A, Vagedes J. How accurate is pulse rate variability as an estimate of heart rate variability? A review on studies comparing photoplethysmographic technology with an electrocardiogram. *Int J Cardiol* 2013;166:15-29
 37. Selvaraj N, Jaryal A, Santhosh J, et al. Assessment of heart rate variability derived from finger-tip photoplethysmography as compared to electrocardiography. *J Med Eng Technol* 2008;32(6):479-84.
 38. Shin H. Ambient temperature effect on pulse rate variability as an alternative to heart rate variability in young adult. *J Clin Monit Comput* 2016;30(6):939-48.
 39. Lee J, Matsumura K, Yamakoshi K, et al. Comparison between red, green and blue light reflection photoplethysmography for heart rate monitoring during motion. *Conf Proc IEEE Eng Med Biol Soc* 2013;2013:1724-7.
 40. Wong JS, Lu WA, Wu KT, et al. A comparative study of pulse rate variability and heart rate variability in healthy subjects. *J Clin Monit Comput* 2012;26(2):107-14.
 41. Pinheiro N, Couceiro R, Henriques J, et al. Can PPG be used for HRV analysis? *Conf Proc IEEE Eng Med Biol Soc* 2016;2016:2945-9.
 42. Lee ES, Lee JS, Joo MC, et al. Accuracy of heart rate measurement using smart phones during treadmill exercise in male patients with ischemic heart disease. *Ann Rehabil Med* 2017;41(1):129-37.
 43. Sun Y, Thakor N. Photoplethysmography revisited: from contact to noncontact, from point to imaging. *IEEE Trans Biomed Eng* 2016;63(3):463-77.
 44. Chen X, Huang Y, Yun F, et al. Effect of changes in sympathovagal balance on the accuracy of heart rate variability obtained from photoplethysmography. *Exp Ther Med* 2015;10(6):2311-8.
 45. Gil E, Orini M, Bailón R, et al. Photoplethysmography pulse rate variability as a surrogate measurement of heart rate variability during non-stationary conditions. *Physiol Meas* 2010;31(9):1271-90.
 46. Tarvainen MP, Niskanen JP, Lipponen JA, et al. Kubios HRV—heart rate variability analysis software. *Comput Methods Programs Biomed* 2014;113(1):210-20.
 47. Bland JM, Altman DG. Statistical methods for assessing agreement between two methods of clinical measurement. *Lancet* 1986;1:307-10.
 48. Qiu Y, Cai Y, Zhu Y, et al. Poincaré plot analysis for pulse interval extracted from non-contact photoplethysmography. *Conf Proc IEEE Eng Med Biol Soc* 2005;2:1964-7.
 49. Chreiteh S, Belhage B, Hoppe K, et al. Sternal pulse rate variability compared with heart rate variability on healthy subjects. *Conf Proc IEEE Eng Med Biol Soc* 2014;2014:3394-7.
 50. Blackford EB, Estep JR. Effects of frame rate and image resolution on pulse rate measured using multiple camera imaging photoplethysmography. In: B. Gimi, & R. C. Molthen, (Eds.). *Proceedings of SPIE 9417, Medical imaging 2015: Biomedical applications in molecular structural and functional imaging*. 2015. p. 1-14.
 51. Iozzia L, Cerina L, Mainardi L. Relationships between heart-rate variability and pulse-rate variability obtained from video-PPG signal using ZCA. *Physiol Meas* 2016;37(11):1934-44.

Возможности переносных приборов для точного измерения вариабельности сердечного ритма. Систематическое обозрение

Константинос Георгиу¹, Андреас В. Ларенцакис², Нехал Н. Камис³, Гада И. Алсунхаибани³, Ясер А. Аласка³, Елиас Дж. Джиалафос⁴

¹ Кафедра биохимии, Медицинский факультет, Национальный Афинский университет им. Каподистрия, Афины, Греция

² Первая клиника хирургической пропедевтики, Медицинский факультет, Национальный Афинский университет им. Каподистрия, Афины, Греция

³ Центр клинических умений и симуляции, Медицинский факультет, Медицинский колледж, Университет короля Сауда, Эр-Рияд, Саудовская Аравия

⁴ Симуляционный центр медицинской физики, Медицинский факультет, Национальный Афинский университет им. Каподистрия, Афины, Греция

Адрес для корреспонденции:

Константинос Георгиу, Кафедра биохимии, Медицинский факультет, Национальный Афинский университет им. Каподистрия, ул. „Микрас Асиас“ 75, Гоуди, 11527 Афины, Греция
E-mail: kongeorgiou@med.uoa.gr
Тел: +306942066216

Дата получения: 27 декабря 2017

Дата приемки: 19 января 2018

Дата онлайн публикации: 31 января 2018

Дата публикации: 30 марта 2018

Ключевые слова: сердечный ритм, вариабельность сердечного ритма, переносные, фотоплетизмография

Образец цитирования:

Georgiou K, Larentzakis AV, Khamis NN, Alsuhaybani GI, Alaska YA, Giialafos EJ. Can wearable devices accurately measure heart rate variability? A systematic review. *Folia Med (Plovdiv)* 2018;60(1):7-20
doi: 10.2478/folmed-2018-0012

Введение: Утверждается, что всё большее количество переносных приборов обеспечивает установление точных, недорогих и легко применимых показателей вариабельности сердечного ритма (ВСР). В основном это обеспечивается приборами переносной фотоплетизмографии (ФПГ) и/или электрокардиографии (ЭКГ) с помощью простых и неинвазивных методов в качестве заместителей золотого стандарта оценки интервала RR с помощью электрокардиограммы. Хотя соответствие между вариабельностью частоты пульса (ВЧП) и ВСР было исследовано в литературе, полученные результаты по-прежнему неубедительны, особенно в отношении использования переносных устройств.

Цель: Целью настоящего систематического обозрения является установление возможностей переносных приборов обеспечивать надёжное и точное измерение классических параметров ВСР как в состоянии покоя, так и во время физической нагрузки.

Материалы и методы: Была использована стратегия сбора данных для нахождения соответствующих статей в базах данных MEDLINE и SCOPUS, а также в Интернете. Найденные 308 статей были рассмотрены для дальнейшей оценки в соответствии с заранее определёнными критериями включения / исключения.

Результаты: Восемнадцать исследований были включены. В шестнадцати из них применялась технология ЭКГ – ВСР, а в двух из них - технология ФПГ – ВЧП. Все они исследовали точность переносных приборов при измерении СР во время отдыха, а только восемь из них во время физической нагрузки. Корреляция между классической ВСР, полученной от ЭКГ, и переносной ВСР варьировалась от очень хорошей до превосходной во время отдыха, но она постепенно снижалась по мере увеличения нагрузки.

Заключение: Переносные приборы могут обеспечить альтернативное решение для измерения ВСР. Тем не менее, необходимы более надёжные исследования в нестационарных условиях с использованием соответствующей методологии в отношении количества участвующих в исследовании субъектов, использованных методов измерения и анализа.