
A SIMPLE, UNIQUE FIELD VENTILATOR FOR LARGE UNGULATES: ANOTHER USE FOR YOUR LEAF BLOWER

Scott B. Citino, DVM, Dipl ACZM,^{1} Mitchell Bush, DVM, Dipl ACZM,² and Oswaldo Rivera³*

¹*White Oak Conservation Center, 581705 White Oak Road, Yulee, FL 32097-2145 USA;*

²*Khao Kheow Open Zoo, 235 Moo7 Bangphra, Siracha, Chonburi 20110, Thailand;*

³*Children's National Medical Center, Washington, D.C. 20010 USA*

Abstract

Field anesthesia of wild ungulates commonly involves the use of opioid-based drug combinations. Hypoxemia, hypercarbia, and acidemia are common problems resulting from respiratory depression seen with opioid-based anesthesia. Oxygen supplementation and/or respiratory support are generally recommended when using opioids. Oxygen supplementation alone reduces hypoxemia but does not improve ventilation and gas exchange.¹ Intermittent, positive-pressure, field ventilation has generally been accomplished in large ungulates by the use of single or multiple, in-series, high flow demand valves using 100% oxygen as the driving gas.³ These systems require pressurized oxygen tanks to be carried into the field which is cumbersome and can be dangerous. Bellows-type ventilators are large, difficult to transport, and also require pressurized driving gases. Blower- or fan-driven ventilators have been used in veterinary medicine to ventilate large domestic species,⁴ including one using an electric vacuum cleaner as the drive source.⁵ A simple, effective ventilator, utilizing either an electric or gas-powered leaf blower as the drive source, was designed to be used as an emergency field ventilator during a giraffe anesthesia study.² This ventilator is made of commonly available parts, has a pop-off valve and pressure gauge to prevent over-inflation, an expiratory valve to provide post-expiratory end pressure (PEEP), and is relatively compact and easy to carry into the field. This device was used to ventilate three captive, sub-adult giraffe ranging in estimated weight from 580 to 700 kg. Inspiratory pressures of 20 cm H₂O were reached quickly (< 2 sec), p_aO₂ values were significantly increased, and p_aCO₂ values were significantly decreased over pre-ventilation values, indicating adequate ventilation efficiency.

ACKNOWLEDGMENTS

This study was funded through a grant from the Morris Animal Foundation (#D03ZO-63).

LITERATURE CITED

1. Bush, M., J. Raath, D. Grobler, and L. Klein. 2004. Severe hypoxaemia in field-anaesthetized white rhinoceros (*Ceratotherium simum*) and effects of using tracheal insufflation of oxygen. *J. S. Afr. Vet. Assoc.* 75:79-84.
2. Citino, S.B., M. Bush, W. Lance, M. Hofmeyr, and D. Grobler. 2006. Use of thiafentanil (A3080), medetomidine, and ketamine for anesthesia of captive and free-ranging giraffe (*Giraffa camelopardalis*). *Proc. Am. Assoc. Zoo Vet.* Pp. 211-213.
3. Horne, W.A., M.N. Tchamba, and M.R. Loomis. 2001. A simple method of providing intermittent positive-pressure ventilation to etorphine-immobilized elephants (*Loxodonta africana*) in the field. *J. Zoo Wildl. Med.* 32:519-522.

-
-
4. Orima, H., M. Tagawa, H. Ejima, M. Shimizu, M. Fujita, and M. Washizu. 1992. Blower-driven artificial respirator for large animals. *J. Vet. Med. Sci.* 54:177-178.
 5. Moore, A.G., J.D. McCrady, J. Vasku, M.E. Geddes, and H.E. Hoff. 1964. A respirator for large animals. *Southwest Vet.* 18:22-24.